

REVIEW ARTICLE

Advances in Robotic-Assisted Surgery: Evaluating Outcomes and Future Directions in Minimally Invasive Procedures

Alen Sam Saji ^a, Sai Han Htun ^b, Lydia Ellen Mao ^{c*}

^a West China School of Medicine, West China Hospital, Sichuan University, Chengdu-610041, China.

^b International School of Medicine, Zhejiang University, Zhejiang-310027, China.

^c Chinese University of Hong Kong, Central Ave, Hong Kong.

Corresponding Author: Lydia Ellen Mao
Chinese University of Hong Kong, Central
Ave, Hong Kong.
Email: lydia200426@gmail.com

Articleinfo

Received: 8 February 2024

Accepted: 10 April 2024

Keywords: Robotic-assisted surgery;
Minimally invasive surgery; Machine
learning; Healthcare access.

How to cite this article: Alen Sam Saji, Sai Han Htun, Lydia Ellen Mao. (2024). Advances in Robotic-Assisted Surgery: Evaluating Outcomes and Future Directions in Minimally Invasive Procedures, 1(1), 28-33 Retrieved from <https://archmedrep.com/index.php/amr/article/view/7>

Abstract

Robotic-assisted surgery (RAS) has transformed the field of minimally invasive surgery by integrating advanced technologies that enhance surgical precision, flexibility, and control. This review delves into the significant technological advances in RAS, particularly focusing on its applications across various surgical disciplines. Innovations such as high-definition 3D vision systems, wristed instruments with multiple degrees of freedom, and sophisticated software providing real-time data and feedback have substantially improved surgical outcomes. These technological advancements have addressed critical challenges such as limited visualization, surgeon fatigue, and imprecise movements associated with traditional laparoscopic surgery. By evaluating the impact of these innovations on patient outcomes, surgical practice, and the broader healthcare system, this review highlights the transformative potential of RAS. Furthermore, it explores the future directions of robotic surgery, including the integration of artificial intelligence, the development of tele-surgery, and the continuous improvement of training programs for surgeons, aiming to push the boundaries of what is possible in surgical care.

1. Introduction

Robotic-assisted surgery (RAS) represents one of the most significant advancements in the field of minimally invasive surgery (MIS) over the past few decades (Dagnino and Kundrat, 2024). Since the introduction of the first robotic surgical system, the da Vinci Surgical System, by Intuitive Surgical in 2000, the technology has rapidly evolved, transforming surgical practices and outcomes across multiple medical disciplines (Tsuda et al., 2015). The primary advantage of RAS lies in its ability to enhance the precision, control, and flexibility of surgeons, leading to improved patient outcomes, reduced recovery times, and minimized surgical trauma (Boehm et al., 2021). MIS, which encompasses techniques like laparoscopy and endoscopy, was initially developed to reduce the invasiveness of traditional open surgeries. These techniques have significantly improved patient care by decreasing postoperative pain, shortening hospital stays, and lowering the risk of infections and complications. However, conventional MIS techniques have limitations, particularly in terms of the range of motion and dexterity of the instruments, as well as the two-dimensional (2D) visualization provided by standard laparoscopic equipment (Vitiello et al., 2012). RAS addresses these limitations by incorporating advanced technologies that provide surgeons

with high-definition three-dimensional (3D) vision, enhanced instrument maneuverability, and ergonomic benefits.

The cornerstone of RAS is the integration of sophisticated robotic systems that allow for greater precision and control during surgical procedures (Klodmann et al., 2021). These systems typically consist of a console where the surgeon sits and controls the robotic instruments, a patient-side cart with robotic arms that hold and manipulate the surgical instruments, and a high-definition 3D vision system that provides the surgeon with a magnified view of the surgical field. The robotic arms translate the surgeon's hand movements into precise micro-movements of the instruments, enabling complex procedures to be performed with greater accuracy and less tissue damage than with traditional surgical techniques (Ibrahim et al., 2012). One of the key technological advancements in RAS is the development of wristed instruments that offer a greater range of motion than the human hand. These instruments can rotate and bend in ways that traditional laparoscopic instruments cannot, allowing for more intricate dissection and suturing (Anderson et al., 2016). Additionally, the high-definition 3D vision systems used in RAS provide surgeons with a clearer and more detailed view of the operative field, enhancing their ability to identify and preserve vital structures. Some systems are also equipped with advanced imaging technologies, such

© The Author(s). 2024 Open Access This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and non-commercial reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated.

as fluorescence imaging, which can help in visualizing blood flow and identifying cancerous tissues (Autorino et al., 2014). RAS has been successfully applied in various surgical specialties, including urology, gynecology, cardiothoracic surgery, and colorectal surgery. In urology, for example, RAS is widely used for procedures such as prostatectomy and nephrectomy, offering improved oncological outcomes and reduced complication rates compared to traditional techniques (Falagario et al., 2020). In gynecology, robotic-assisted hysterectomy and myomectomy have become increasingly common, providing patients with faster recovery times and less postoperative pain (Lonnerfors, 2018). Similarly, in cardiothoracic surgery, RAS has enabled minimally invasive approaches to complex procedures such as mitral valve repair and coronary artery bypass grafting, reducing the need for sternotomies and improving patient recovery (Marin Cuartas et al., 2017).

Despite the numerous advantages of RAS, several challenges remain. The high cost of robotic systems and the associated maintenance and training expenses can be prohibitive for many healthcare institutions, limiting the widespread adoption of this technology. Additionally, there is a need for standardized training programs and credentialing processes to ensure that surgeons are proficient in using robotic systems and can deliver optimal patient outcomes. As the technology continues to evolve, ongoing research and clinical trials are essential to evaluate the long-term benefits and potential risks of RAS and to refine the techniques and protocols used in various surgical specialties.

2. Technological Advances in RAS

2.1. Enhanced Imaging Modalities

Modern robotic surgical systems integrate advanced imaging modalities such as intraoperative CT scans, MRI fusion, and optical coherence tomography (OCT). These technologies provide detailed, real-time visualization of the surgical field with high resolution, enabling surgeons to navigate complex anatomical structures more precisely. For example, OCT can visualize tissue microstructures in real-time during ophthalmic surgeries, while intraoperative MRI facilitates accurate tumor resections in neurosurgery (Assayag et al., 2013).

2.2. Robotics in Orthopedic Surgery

Robotic-assisted orthopedic surgery has transformed procedures like total knee arthroplasty (TKA) and total hip arthroplasty (THA). Robotic systems use preoperative imaging data to create a patient-specific 3D model, enabling precise bone resections and optimal implant positioning (Chen et al., 2016). This customization improves joint alignment, stability, and functional outcomes for patients, reducing the risk of complications such as implant loosening and leg length discrepancy.

2.3. Integration of Augmented Reality (AR)

Augmented reality overlays computer-generated images onto the surgeon's view of the patient, enhancing spatial orientation and procedural planning (Gao et

al., 2021). AR can superimpose anatomical structures, preoperative imaging data, and instrument tracking information directly onto the surgeon's field of view through head-mounted displays or surgical microscopes. This technology assists in complex surgeries by providing real-time guidance, improving accuracy, and reducing surgical errors.

2.4. Advancements in Haptic Feedback

Haptic feedback systems in robotic surgery simulate the tactile sensation of touch and force feedback to the surgeon's hands (Abiri et al., 2019). By transmitting forces exerted on robotic instruments back to the surgeon, haptic feedback enhances tactile perception and dexterity during delicate maneuvers. This capability is crucial in procedures like tissue dissection, knot tying, and suturing, where precise force control is essential to avoid tissue damage and ensure optimal surgical outcomes.

2.5. Remote Telesurgery and Collaborative Robotics

Remote telesurgery allows surgeons to perform procedures from distant locations using robotic systems equipped with high-speed internet connectivity and low-latency communication channels (Navarro et al., 2022). Collaborative robotics enables multiple surgeons to control different robotic arms simultaneously during complex surgeries, fostering teamwork and expertise exchange. These capabilities enhance surgical access in underserved areas and facilitate specialized surgical care delivery worldwide (Figure 1 & Table 1).

3. Applications of RAS

3.1. Minimally Invasive Gynecological Surgery

Robotic systems are employed in gynecological procedures such as hysterectomy, myomectomy (fibroid removal), and ovarian cystectomy. The minimally invasive approach reduces post-operative pain, shortens hospital stays, and minimizes scarring compared to traditional open surgeries. Robotic precision enables intricate maneuvers in the confined pelvic cavity, preserving surrounding organs and enhancing patient recovery (Kim et al., 2017).

3.2. Head and Neck Surgery

RAS is utilized in head and neck procedures including transoral robotic surgery (TORS) for tumors of the throat, tongue, and larynx. The articulated robotic arms maneuver through narrow anatomical spaces with enhanced reach and flexibility, enabling precise tumor resection while preserving vital structures such as nerves and blood vessels. TORS reduces post-operative complications, speech impairment, and swallowing difficulties compared to conventional approaches.

3.3. Complex Colorectal Surgeries

Robotic platforms facilitate complex colorectal surgeries such as low anterior resection for rectal cancer and ileal pouch-anal anastomosis (IPAA) for ulcerative colitis (Morelli et al., 2015). The dexterity of robotic instruments allows for precise dissection in the narrow pelvis, optimal

anastomotic techniques, and reduced risk of complications like anastomotic leakage. Enhanced visualization and instrument articulation improve surgical outcomes, bowel function preservation, and patient quality of life.

3.4. Thoracic Surgery

Robotic-assisted thoracic surgery (RATS) is employed in procedures such as lobectomy for lung cancer and thymectomy for thymoma (Mattioni et al., 2022). The minimally invasive approach reduces post-operative pain, respiratory complications, and hospitalization duration compared to thoracotomy. Robotic systems enable precise dissection around delicate structures in the chest cavity, promoting lung preservation and functional outcomes in patients undergoing oncological resections (Table 2).

4. Clinical Outcomes of RAS

The clinical outcomes of RAS have been extensively studied, with numerous benefits observed compared to traditional surgical methods. RAS results in smaller incisions, leading to less pain, shorter hospital stays, and quicker recovery times. Enhanced precision during surgery reduces intraoperative blood loss and the need for transfusions. Patients undergoing RAS typically experience fewer complications and faster returns to normal activities (Ahmad et al., 2017). The enhanced precision of robotic systems reduces the likelihood of errors and improves surgical outcomes. Consistent, reproducible results are particularly valuable in complex procedures. The accuracy and control provided by RAS enable surgeons to perform delicate tasks with confidence. In oncological surgeries, RAS allows for more precise tumor resections, potentially improving long-term survival rates. Enhanced visualization and dexterity enable thorough lymph node dissections, critical for accurate staging and treatment. RAS is used in various specialties, including urology, gynecology, cardiothoracic, and colorectal surgery. The technology facilitates complex procedures

that are challenging or impossible with conventional laparoscopic methods. As a result, RAS has expanded the range of minimally invasive procedures available to patients.

5. Challenges and Limitations of RAS

Despite the many advantages of RAS, there are several challenges and limitations that need to be addressed. The high cost of robotic systems and the associated maintenance expenses can be a significant barrier for many healthcare institutions. The initial investment in robotic systems can be substantial, making it challenging for smaller hospitals and clinics to adopt this technology. Additionally, the cost of disposable instruments and ongoing maintenance can add to the financial burden. Learning to use robotic systems effectively requires specialized training and experience. Surgeons must undergo extensive training to master the skills necessary for RAS. The learning curve can be steep, and proficiency may take time to achieve. This can limit the widespread adoption of RAS, particularly in areas with limited access to training programs. As RAS becomes more widespread, ethical considerations surrounding its use must be addressed (Schreyer et al., 2022). Issues such as the equitable distribution of technology, informed consent, and the potential for overuse of RAS in cases where traditional methods may be equally effective need careful consideration.

6. Future Directions in RAS

The future of RAS holds promise for even greater advancements and applications across various surgical disciplines. Continued technological advancements will likely lead to the development of more sophisticated robotic systems. These systems may offer improved haptic feedback, enhanced imaging capabilities, and greater automation. The integration of artificial intelligence (AI) and machine learning could further enhance the precision and efficiency of RAS (Moglia et al., 2021).




Figure 1: Innovations in Robotic assisted surgery technologies

Table 1: Innovations in Robotic-Assisted Surgery Technologies

Technology	Description	Application	Reference
Enhanced Imaging Modalities	Integration of intraoperative CT scans, MRI fusion, and optical coherence tomography (OCT)	Neurosurgery, ophthalmic surgery	(Schuppper et al., 2021)
Robotics in Orthopedics	Patient-specific 3D models for precise bone resections and implant positioning	Total knee arthroplasty, hip arthroplasty	(Jacofsky and Allen, 2016)
Augmented Reality (AR)	Overlay of anatomical structures and instrument tracking data onto the surgeon's view	Complex surgeries, procedural guidance	(Liu et al., 2019)
Haptic Feedback Systems	Simulation of tactile sensation and force feedback to enhance surgical precision	Tissue dissection, suturing	(Zhou et al., 2012)
Remote Telesurgery	High-speed internet connectivity for remote surgery execution	Global surgical access	(Haidegger et al., 2011)

Table 2: Clinical Applications of Robotic-Assisted Surgery

Medical Specialty	Surgical Procedure	Surgical Description	Reference
Gynecological Surgery	Hysterectomy, myomectomy	Minimally invasive approach reduces post-operative pain and scarring	(Loddo et al., 2022)
Head and Neck Surgery	Transoral robotic surgery (TORS)	Precise tumor resection while preserving nerves and blood vessels	(Lawson et al., 2011)
Colorectal Surgery	Low anterior resection, IPAA	Enhanced precision in pelvic dissection, reduced complications like anastomotic leakage	(Capolupo et al., 2021)
Thoracic Surgery	Lobectomy, thymectomy	Minimally invasive approach improves lung preservation, reduces respiratory complications	(Yang et al., 2016)

Tele-surgery, where surgeons operate remotely using robotic systems, has the potential to revolutionize surgical care. This technology could provide access to specialized surgical expertise in remote or underserved areas, improving patient outcomes and reducing disparities in healthcare access. As the field of RAS evolves, it is essential to focus on developing and refining training programs for surgeons. Simulation-based training and virtual reality technologies could provide valuable tools for enhancing surgical skills and reducing the learning curve associated with RAS (Sinha et al., 2023). The ongoing integration of AI and machine learning in RAS could enhance decision-making and improve surgical outcomes. AI algorithms could assist in identifying optimal surgical approaches, predicting complications, and providing real-time feedback during procedures.

7. Conclusion

RAS has significantly advanced the field of MIS, offering

numerous benefits over traditional methods. Technological advancements have enhanced surgical precision, improved patient outcomes, and expanded the range of procedures that can be performed minimally invasively. However, challenges such as high costs, the need for specialized training, and ethical considerations must be addressed to ensure the widespread adoption and equitable distribution of RAS. The future of RAS holds promise for even greater advancements, with the potential to revolutionize surgical care and improve patient outcomes further. Continued innovation and investment in this field will be essential to realizing the full potential of RAS

Declarations

Ethics approval statement

No ethical approval was required for the current study as it did not deal with any human or animal samples.

Consent to participate

Not applicable

Consent to publish

Not applicable

Data Availability Statement

The data are available from the corresponding author upon reasonable request

Competing Interests

The authors declare that they have no conflict of interest

Funding

Not Applicable

Author contribution

A.S.S and S.H.H: investigation, formal analysis, writing original draft. L.E.M: conceptualization, writing original draft, and supervision.

Acknowledgements

Not Applicable

Reference

1. Abiri, A., Juo, Y.-Y., Tao, A., Askari, S.J., Pensa, J., Bisley, J.W., Dutson, E.P., Grundfest, W.S., 2019. Artificial palpation in robotic surgery using haptic feedback. *Surg. Endosc.* 33, 1252–1259. <https://doi.org/10.1007/s00464-018-6405-8>
2. Ahmad, A., Ahmad, Z.F., Carleton, J.D., Agarwala, A., 2017. Robotic surgery: current perceptions and the clinical evidence. *Surg. Endosc.* 31, 255–263. <https://doi.org/10.1007/s00464-016-4966-y>
3. Anderson, P.L., Lathrop, R.A., Webster III, R.J., 2016. Robot-like dexterity without computers and motors: a review of hand-held laparoscopic instruments with wrist-like tip articulation. *Expert Rev. Med. Devices* 13, 661–672. <https://doi.org/10.1586/17434440.2016.1146585>
4. Assayag, O., Grieve, K., Devaux, B., Harms, F., Pallud, J., Chretien, F., Boccara, C., Varlet, P., 2013. Imaging of non-tumorous and tumorous human brain tissues with full-field optical coherence tomography. *NeuroImage Clin.* 2, 549–557. <https://doi.org/10.1016/j.nicl.2013.04.005>
5. Autorino, R., Zargar, H., White, W.M., Novara, G., Annino, F., Perdonà, S., De Angelis, M., Mottrie, A., Porpiglia, F., Kaouk, J.H., 2014. Current Applications of Near-infrared Fluorescence Imaging in Robotic Urologic Surgery: A Systematic Review and Critical Analysis of the Literature. *Urology* 84, 751–759. <https://doi.org/10.1016/j.urology.2014.05.059>
6. Boehm, F., Graesslin, R., Theodoraki, M.-N., Schild, L., Greve, J., Hoffmann, T.K., Schuler, P.J., 2021. Current Advances in Robotics for Head and Neck Surgery—A Systematic Review. *Cancers (Basel)*. 13, 1398. <https://doi.org/10.3390/cancers13061398>
7. Capolupo, G.T., Carannante, F., Mascianà, G., Lauricella, S., Mazzotta, E., Caricato, M., 2021. Transanal proctocolectomy and ileal pouch-anal anastomosis (TaIPAA) for ulcerative colitis: medium term functional outcomes in a single centre. *BMC Surg.* 21, 17. <https://doi.org/10.1186/s12893-020-01007-z>
8. Chen, X., Xu, L., Wang, Y., Hao, Y., Wang, L., 2016. Image-guided installation of 3D-printed patient-specific implant and its application in pelvic tumor resection and reconstruction surgery. *Comput. Methods Programs Biomed.* 125, 66–78. <https://doi.org/10.1016/j.cmpb.2015.10.020>
9. Dagnino, G., Kundrat, D., 2024. Robot-assistive minimally invasive surgery: trends and future directions. *Int. J. Intell. Robot. Appl.* <https://doi.org/10.1007/s41315-024-00341-2>
10. Falagario, U., Veccia, A., Weprin, S., Albuquerque, E. V., Nahas, W.C., Carrieri, G., Pansadoro, V., Hampton, L.J., Porpiglia, F., Autorino, R., 2020. Robotic-assisted surgery for the treatment of urologic cancers: recent advances. *Expert Rev. Med. Devices* 17, 579–590. <https://doi.org/10.1080/17434440.2020.1762487>
11. Gao, Y., Zhao, Y., Xie, L., Zheng, G., 2021. A Projector-Based Augmented Reality Navigation System for Computer-Assisted Surgery. *Sensors* 21, 2931. <https://doi.org/10.3390/s21092931>
12. Haidegger, T., Sándor, J., Benyó, Z., 2011. Surgery in space: the future of robotic telesurgery. *Surg. Endosc.* 25, 681–690. <https://doi.org/10.1007/s00464-010-1243-3>
13. Ibrahim, A.E., Sarhane, K.A., Baroud, J.S., Atiyeh, B.S., 2012. Robotics in plastic surgery, a review. *Eur. J. Plast. Surg.* 35, 571–578. <https://doi.org/10.1007/s00238-012-0737-8>
14. Jacofsky, D.J., Allen, M., 2016. Robotics in Arthroplasty: A Comprehensive Review. *J. Arthroplasty* 31, 2353–2363. <https://doi.org/10.1016/j.arth.2016.05.026>
15. Kim, S., Luu, T.H., Larena, N., Falcone, T., 2017. Role of robotic surgery in treating fibroids and benign uterine mass. *Best Pract. Res. Clin. Obstet. Gynaecol.* 45, 48–59. <https://doi.org/10.1016/j.bpobgyn.2017.04.004>
16. Klodmann, J., Schlenk, C., Hellings-Kuß, A., Bahls, T., Unterhinninghofen, R., Albu-Schäffer, A., Hirzinger, G., 2021. An Introduction to Robotically Assisted Surgical Systems: Current Developments and Focus Areas of Research. *Curr. Robot. Reports* 2, 321–332. <https://doi.org/10.1007/s43154-021-00064-3>
17. Lawson, G., Matar, N., Remacle, M., Jamart, J., Bachy, V., 2011. Transoral robotic surgery for the management of head and neck tumors: learning curve. *Eur. Arch. Oto-Rhino-Laryngology* 268, 1795–1801. <https://doi.org/10.1007/s00405-011-1537-7>
18. Liu, J., Al'Aref, S.J., Singh, G., Caprio, A., Moghadam, A.A.A., Jang, S.-J., Wong, S.C., Min, J.K., Dunham, S., Mosadegh, B., 2019. An augmented reality system for image guidance of transcatheter procedures for structural heart disease. *PLoS One* 14, e0219174. <https://doi.org/10.1371/journal.pone.0219174>
19. Loddo, A., Djokovic, D., Drizi, A., De Vree, B.P., Sedrati, A., van Herendael, B.J., 2022. Hysteroscopic myomectomy: The guidelines of the International Society for Gynecologic Endoscopy (ISGE). *Eur. J. Obstet. Gynecol. Reprod. Biol.* 268, 121–128. <https://doi.org/10.1016/j.ejogrb.2021.11.434>
20. Lonnerfors, C., 2018. Robot-assisted myomectomy. *Best Pract. Res. Clin. Obstet. Gynaecol.* 46, 113–119. <https://doi.org/10.1016/j.bpobgyn.2017.09.005>
21. Marin Cuartas, M., Javadikasgari, H., Pfannmueller, B., Seeburger, J., Gillinov, A.M., Suri, R.M., Borger, M.A.,

2017. Mitral valve repair: Robotic and other minimally invasive approaches. *Prog. Cardiovasc. Dis.* 60, 394–404. <https://doi.org/10.1016/j.pcad.2017.11.002>
22. Mattioni, G., Palleschi, A., Mendogni, P., Tosi, D., 2022. Approaches and outcomes of Robotic-Assisted Thoracic Surgery (RATS) for lung cancer: a narrative review. *J. Robot. Surg.* 17, 797–809. <https://doi.org/10.1007/s11701-022-01512-8>
23. Moglia, A., Georgiou, K., Georgiou, E., Satava, R.M., Cuschieri, A., 2021. A systematic review on artificial intelligence in robot-assisted surgery. *Int. J. Surg.* 95, 106151. <https://doi.org/10.1016/j.ijsu.2021.106151>
24. Morelli, L., Guadagni, S., Mariniello, M.D., Furbetta, N., Pisano, R., D'Isidoro, C., Caprili, G., Marciano, E., Di Candio, G., Boggi, U., Mosca, F., 2015. Hand-assisted hybrid laparoscopic-robotic total proctocolectomy with ileal pouch-anal anastomosis. *Langenbeck's Arch. Surg.* 400, 741–748. <https://doi.org/10.1007/s00423-015-1331-x>
25. Navarro, E.M., Ramos Álvarez, A.N., Soler Anguiano, F.I., 2022. A new telesurgery generation supported by 5G technology: benefits and future trends. *Procedia Comput. Sci.* 200, 31–38. <https://doi.org/10.1016/j.procs.2022.01.202>
26. Schreyer, J., Koch, A., Herlemann, A., Becker, A., Schlenker, B., Catchpole, K., Weigl, M., 2022. RAS-NOTECHS: validity and reliability of a tool for measuring non-technical skills in robotic-assisted surgery settings. *Surg. Endosc.* 36, 1916–1926. <https://doi.org/10.1007/s00464-021-08474-2>
27. Schuppper, A.J., Rao, M., Mohammadi, N., Baron, R., Lee, J.Y.K., Acerbi, F., Hadjipanayis, C.G., 2021. Fluorescence-Guided Surgery: A Review on Timing and Use in Brain Tumor Surgery. *Front. Neurol.* 12. <https://doi.org/10.3389/fneur.2021.682151>
28. Sinha, A., West, A., Vasdev, N., Sooriakumaran, P., Rane, A., Dasgupta, P., McKirdy, M., 2023. Current practises and the future of robotic surgical training. *Surg.* 21, 314–322. <https://doi.org/10.1016/j.surge.2023.02.006>
29. Tsuda, S., Oleynikov, D., Gould, J., Azagury, D., Sandler, B., Hutter, M., Ross, S., Haas, E., Brody, F., Satava, R., 2015. SAGES TAVAC safety and effectiveness analysis: da Vinci® Surgical System (Intuitive Surgical, Sunnyvale, CA). *Surg. Endosc.* 29, 2873–2884. <https://doi.org/10.1007/s00464-015-4428-y>
30. Vitiello, V., Kwok, K.-W., Yang, G.-Z., 2012. Introduction to robot-assisted minimally invasive surgery (MIS), in: *Medical Robotics*. Elsevier, pp. 1-P1. <https://doi.org/10.1533/9780857097392.1>
31. Yang, Y., Dong, J., Huang, Y., 2016. Thoracoscopic thymectomy versus open thymectomy for the treatment of thymoma: A meta-analysis. *Eur. J. Surg. Oncol.* 42, 1720–1728. <https://doi.org/10.1016/j.ejso.2016.03.029>
32. Zhou, M., Tse, S., Derevianko, A., Jones, D.B., Schweitzberg, S.D., Cao, C.G.L., 2012. Effect of haptic feedback in laparoscopic surgery skill acquisition. *Surg. Endosc.* 26, 1128–1134. <https://doi.org/10.1007/s00464-011-2011-8>